



Government of West Bengal  
Office of the Director  
Institute of Post Graduate Medical Education & Research  
244 A. J C Bose Road, Kolkata -700 020  
Email :director.ipgmer@gmail.com. Phone : 033-2204-1101

Memo No: Inst/ 1869


Date: 27.03.2024

## CORRIGENDUM NOTICE

In reference to the pre bid meeting (2<sup>nd</sup> call) dated on 21.03.2024, the following amendments are being made in e-tender notice invited vide Tender Reference No- **WBHF/Inst/1597** dated 13.03.2024 of I.P.G.M.E. & R- S.S.K.M Hospital, Kolkata. (Tender ID 2024\_HFW\_684727\_1)


(Fresh specifications are uploaded herewith).

The interested bidders may see the eProcurement System of Government of West Bengal.

  
Director  
IPGMER, Kolkata – 20  
Director  
I.P.G.M.E. & R.  
Kolkata-700020

## Technical Specification of Repetitive Transcranial Magnetic Stimulator (rTMS) 100Hz

1. It should be a highly effective non-invasive repetitive biphasic magnetic stimulator for bilateral cortical or peripheral and central nerve pathways stimulation.
2. The system should be able to deliver output power as per the following
  - a. Should deliver 100% power at 41-42 Hz
  - b. Should deliver 50% output at 96-100 Hz
3. The system should have
  - a. Integrated color touch screen display.
  - b. Integrated 2 channel Motor Evoked Potential.
  - c. *Should be programmable to deliver Single pulse or Repetitive train of sessions.*
  - d. Theta Burst Protocol.
4. The pulse train interval should be user selectable with 0.1s increment and pulse width of 400 us with maximum upper frequency of 100Hz.
5. The system should be CE / ISO certified.
6. The system should be supplied with following Accessories:
  - a. 70 to 100 mm double Air film coil- 1 no. The coil should have integrated air circulation system which should deliver up to 3000 pulses at 120% motor threshold at 10Hz stimulation (Biphasic waveform) over a 30-minute period and remains within allowed temperature limits.
  - b. 70 to 100 mm Double Air Film Sham Coil - 1 no.
7. Should be supplied with patient comfortable chair
8. 1 channel Bluetooth battery operated near infrared spectroscopy for monitoring oxygenation and deoxygenation.



**Prof.(Dr.) Rajesh Pramanik**  
Regn.No.- WBMC 54594  
Professor and H.O.D.  
Dept.of Physical Medicine &  
Rehabilitation  
IPGME &R and SSKM Hospital, Kolkata - 20



**Dr. Firdaus Kamal**  
Professor of PMR  
IPGME&R-SSKM Hospital  
Kolkata-700020



**Dr. Tuhin Ray**  
Regn.No.- WBMC 66159  
Assistant Professor  
Dept.of Physical Medicine &  
Rehabilitation  
IPGME &R and SSKM Hospital, Kolkata- 20